


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-423-PN-PS-05*

S.N.: *7269*

Date: *MAY 23 1995*



- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
4 - 100 NS
- c) Rise time:
 ≤ 200 PS
- d) Fall time:
 ≤ 4 NS
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

120/240 VOLTS
50-60 HZ
A

60 db ATTEN

100 VOLTS/DIV

TOP: 10 NS/DIV

BOT: 200 PS/DIV
(RISE TIME)