


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14-5A-PS-N*

S.N.: *7271*

Date: *JUNE 29 95*



- a) Output signal amplitude:
0 V - 20 V
- b) Pulse width:
200 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
100 ± 10% V
50 - 60 MHz

40 dB 19550V
10 V0 03 101W
PRF = 1.0 MHz

100 ± 10% V
50 - 60 MHz
[Signature]