


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-N-10-P*

S.N.: *7272*

Date: *MAY 29 1995*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS / DIV
PRF ≈ 100 kHz

120 / 240 V
50 - 60 Hz
[Signature]