


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-AV-1-C-PN*

S.N.: *7273*

Date: *APRIL 18 1995*



60 db ATTEN.
100 VOLTS/DIV
PRF = 5 KHz

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
3 TO 100 NS
- c) Rise time:
 ≤ 1 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS

*(Checked at
100 \pm 10 VOLTS)*
[Signature]