


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMR-1-P-ET-W-EO-TRF-ECL-M1*

S.N.: *7280*

Date: *MAY 15 1995*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
5 TO 50 NS
(20% MAX DUTY CYCLE)
- c) Rise time:
≤ 150 PS OR ≤ 1 NS
- d) Fall time:
≤ 150 PS OR ≤ 1 NS
- e) PRF:
0 TO 10 MHz
(20% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:

+24V ± 1.0V

300 mA (max)

[Signature]

40 db ATTEN
5 V/283/DIV
TOP: 2 NS/DIV) SAME
BOT: 20 NS/DIV) PRF

TRR + TR Low