


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-T-P-0P1

S.N.: 7294

Date: JULY 20 1995



- a) Output signal amplitude:
0 TO +350 Volts
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
 ≤ 2 NS
- d) Fall time:
 ≤ 3 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz
[Signature]

70 dB ATTEN
~ 320 VOLTS/DIV
PRF = 3.0 KHz
LOCAL MODE