


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV0-5-C-PW
 S.N.: 7299
 Date: MAY 18 1995



- a) Output signal amplitude:
0 to +28 mV
- b) Pulse width:
6 to 100 ns
- c) Rise time:
≤ 3 ns
- d) Fall time:
≤ 4 ns
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

AV0-5-TT output
 to 30 kHz
 100 V/V / DIV.
 PRF ≈ 2 kHz

[Signature]