


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-7AV-1-e-N

S.N.: 7300

Date: MAY 25 1995



- a) Output signal amplitude:
0 to 100 VOLTS
- b) Pulse width:
3 to 100 NS
- c) Rise time:
51 NS
- d) Fall time:
52 NS
- e) PRF:
0 to 5 KHz
- f) Jitter, stability:
OK

- g) Prime power:
100 / 200 VOLTS
(± 10%)
50-60 Hz

60 db ATTN.: 100 VOLTS/DIV
TOP 10 NS/DIV

BOI 1 NS/DIV
(RISE TIME)

PRF = 5 KHz