


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-155A-C-P*
 S.N.: *7303*
 Date: *MAY 17 1995*

- a) Output signal amplitude:
PULSE: 0 TO +200 mA
DC: 0 TO +200 mA
- b) Pulse width:
20 NS TO 2.0 μS
- c) Rise time: $\leq 10 \text{ NS}$
- d) Fall time: $\leq 10 \text{ NS}$
- e) PRF: *6 TO 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:



120/240 V
50-60 Hz

TOP: 50 NS/DIV, 1V/DIV
BOT: 500 NS/DIV, 1V/DIV

