


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MMP-2-C-PN*

S.N.: *7305*

Date: *JUNE 19 1995*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/DIV
PRF = 1.0 MHz

TOP: 5 NS/DIV

BOT: 500 PS/DIV
(RISE TIME)

120/240 V
50-60 Hz
AK