


PULSE GENERATOR
PERFORMANCE CHECK

Model: AM-7B-C-P

S.N.: 7308

Date: JUNE 21 1995



- a) Output signal amplitude:
0 TO +700 VOLTS TO
- b) Pulse width: $V_L \approx 50 \text{ ns}$
0.1 TO 100 μs
(0.25% MAX DUTY CYCLE)
- c) Rise time:
 $\leq 50 \text{ ns}$
- d) Fall time:
 $\leq 50 \text{ ns}$
- e) PRF: 0 TO 10 KHz
(0.25% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 Hz

100 VOLTS/DIV
 $V_L \approx 50 \text{ ns}$
PRF = 50 Hz