


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MR-B3-C-H51 PN*

S.N.: *7323*

Date: *JULY 20 1995*



- a) Output signal amplitude:  
*0 TO ± 200 VOLTS*
- b) Pulse width: (*TO CL ≤ 200 pF*)  
*20 TO 200 NS*
- c) Rise time: (*10% MAX DUTY CYCLE*)  
*≤ 5 NS*


- d) Fall time:  
*≤ 10 NS*

- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*10%, MAX DUTY CYCLE*  
*OK*

- g) Prime power:  
*120/240 VOLTS*  
*50-60 Hz*

*[Signature]*

*(A) Pout CL = 200 pF  
PRF = 100 kHz*



*(B) Nout CL = 200 pF  
PRF = 100 kHz*