


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MH-141-PN-SSC1*

S.N.: *7326*

Date: *JULY 17 1995*



- a) Output signal amplitude:
± 100 VOLTS.
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
15 VOLTS
250 mA MAX

TOP: P_{out}
BOT: N_{out}
60 db ATTEN
100 VOLTS/PH
PRF = 100 KHz