


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUH-HV1-PIN-SSCI*
S.N.: *7327*
Date: *JULY 25 1995*



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0N
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

TOP: P_{out}
BOT: N_{out}
60 db ATTEN
100 VOLTS PW
PRF = 30 KHz

+15 VOLTS
250 mA