


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PH-401-PN-5C1*

S.N.: *7328*

Date: *JUN 26 1995*



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

TOP: P out

BOT: N out

60 db ATTN

100 VOLTS/DIV

PRF = 30 KHz

± 15 VOLTS
250 mA MAX

[Signature]