


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH - HVI - PN - SSC1

S.N.: 7329

Date: JULY 31, 1995



a) Output signal amplitude:

$\pm$  100 VOLTS

b) Pulse width:

0.6 TO 2.0 NS

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 KHZ

f) Jitter, stability:

0K

g) Prime power:

+ 15 VOLTS

250 mA MAX

TOP: P OUT

BOT: N OUT

60 db ATTN

100 VOLTS 1 DIV

PRF = 30 KHZ