


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HVI-PN-SSC1

S.N.: 7332

Date: JULY 31, 1995



- a) Output signal amplitude:
 ± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VOLTS
250 mA MAX

TOP : P OUT
BOT : N OUT
60 db ATTN
100 VOLTS / DIV
PRF = 30 KHZ