


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-HVI-PN-SSC1

S.N.: 7334

Date: JULY 31, 1995



- a) Output signal amplitude:  
± 100 VOLTS
- b) Pulse width:  
0.6 TO 2.0 NS
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 100 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+ 15 VOLTS  
250 mA MAX

TOP: POUT  
BOT: NOUT  
60 db ATTN  
100 VOLTS / DIV  
PRF = 30 KHZ