


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HVI-PN-SSCI

S.N.: 7335

Date: JULY 31/95



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VOLTS
250 mA MAX

TOP: P out

BOT: N out

60 db ATTN

100 VOLTS / DIV

PRF = 30 KHZ