


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-AN-1-P*

S.N.: *7349*

Date: *AUG 10 1995*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
- f) Jitter, stability:
- g) Prime power:
OK

*TRF: 10 VOLTS/DIV
1 NS/DIV*

*BOT: 10 VOLTS/DIV
100 PS/DIV*

*PRF = 300 KHz
(40 dB ATTEN)*

*+15 VDC
250 mA (MAX)*