


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-3AH-C-P


S.N.: 7354

Date: Aug 1 1995



- a) Output signal amplitude:
0 TO +450 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/248 VOLTS
50-60 Hz

Ⓐ 70 db ATTN
= 320 VOLTS/DIV



Ⓑ AS Ⓐ BUT 2 NS/DIV
(RISE TIME)

[Handwritten signature]