


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AN-15-PS-0T-P*

S.N.: *7375*

Date: *AUG 24 1985*



- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
*≤ 60 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 db ATTEN*

*5 VOLTS/DIV*

*TOP 1 NS/DIV*

*Bot 100 PS/DIV  
(RISE TIME)*

*100V ± 10%*  
*50 → 60 MHz*