


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMH-HM1-PS-P*

S.N.: *7383*

Date: *Aug 16 1995*



60 db ATTEN
100 VOLTS/DIV
PRF = 100 KHz

a) Output signal amplitude:

0 TO 700 VOLTS

b) Pulse width:

≤ 1.0 NS

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

[Signature]