


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-3-PW-C-PN-PWT-AT-OP2


S.N.: 7389

Date: Aug 29 1995



- a) Output signal amplitude:
0 to ± 200 Volts
- b) Pulse width: $R_L \geq 50\Omega$
0.1 to 100 μ s
- c) Rise time:
 ≤ 10 ns
- d) Fall time:
 ≤ 10 ns
- e) PRF: 0 to 10 kHz
- f) Jitter, stability:
OK
- g) Prime power:
170/240 Volts
50-60 Hz

Ⓐ P_{out} . $R_L = 50\Omega$
PRF = 5 kHz



Ⓑ N_{out} . $R_L = 50\Omega$
PRF = 100 Hz