


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD-5D-PS-P-0P1*

S.N.: *7391*

Date: *JUL 27 1995*



- a) Output signal amplitude:
0 to +30 AMP
- b) Pulse width: *($R_L = 3 \Omega$)*
0.5 to 5.0 μ s
- c) Rise time:
 ≤ 40 ns
- d) Fall time:
 ≤ 40 ns
- e) PRF: *6 to 300 Hz*
- f) Jitter, stability:
OK
- g) Prime power:

120/240 WATT
50-60 Hz
[Signature]

