

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM11-4-C-P-NRC2*

S.N.: *7392*

Date: *OCT 6 1996*

a) Output signal amplitude:

*0 TO +100 VOLTS*

b) Pulse width:

*2 TO 4 NS*

c) Rise time:

*≤ 1 NS*

d) Fall time:

*≤ 1.5 NS*

e) PRF:

*0 TO 1.0 MHz*


f) Jitter, stability:

*OK*

g) Prime power:

*120/240 VOLTS*

*50-60 Hz.*



*60 db ATTEN*

*100 VOLTS/DIV*

*TOP: 5 NS/DIV*

*BOT: 1 NS/DIV*

*PRF ≈ 100 KHz*