


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MU-155A-C-N-NYMA1*

S.N.: *7399*


Date: *SEPT 28 1995*



- a) Output signal amplitude:  
*0 TO -100 mA*
- b) Pulse width:  
*10 NS TO 1.0 MS*
- c) Rise time:  
*≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF:  
*1 Hz TO 10 kHz*
- f) Jitter, stability:

*(A)*  
*TOLD 9150*  
*PLUG IN*  
*SOCKET.*  
*100 mA / DIV*  
*20 NS / DIV (MONITOR OUT)*  
*TO 1 MEB*

*OK*  
g) Prime power:  
*120 / 240 V*  
*50-60 Hz*



*(B) HAS A BUT 100 NS / DIV*  
*(RISE TIME)*