


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AV-14V3-F*

S.N.: *7408*

Date: *APR 21 1995*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 - 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

50 db ATTEN

≈ 32 VOLTS/DIV

PRF = 300 KHz,

+15 VDC

200 mA (MAX)

