


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-9A-CN*


S.N.: *7415*

Date: *SEPT 12 1995*



- a) Output signal amplitude:  
*0 TO - 200 mA*
- b) Pulse width:  
*0.4 TO 4.0ns*
- c) Rise time:  
*≤ 200 ps*
- d) Fall time:  
*≤ 200 ps*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*(A) MAINFRAME OUT TO  
50 ~ 40 db ATTEN  
10 VOUTH / DIV*



*120 / 240 V  
50 60 Hz  
[Signature]*

*(B) AIR-SI ME OUT  
TO 20 db  
PRF = 100 kHz*