


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-3A-C-P

S.N.: 7417

Date: SEPT 21 1995



- a) Output signal amplitude:
0 TO +300 VOLTS
- b) Pulse width:
5 TO 100ns
- c) Rise time:
≤ 600 ps
- d) Fall time:
≤ 3 ns
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

TOP: 10 NS/DIV,
320 VOLTS/DIV (70 db.)

120/240 V
50⁺-60 kHz.

BOT: AS TOP BUT
1 NS/DIV
(RISE TIME)

PRF ≈ 2.5 kHz