


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-64-C-N*

S.N.: *7421*

Date: *JAN 12 1996*



- a) Output signal amplitude:
0 TO -200 VOLTS
- b) Pulse width: ($R_L \geq 100K$)
0.5 μ S TO 0.5 MMS
- c) Rise time: (CONTROLLED BY PRF)
 $\leq 10 NS$
- d) Fall time:
 $\leq 10 NS$
- e) PRF: *1 KHZ TO 1 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V 50/60 Hz

50-60 Hz

$R_L = 100K$

PRF $\approx 10 KHz$