


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV9-FD2-PS-IP-EP*

S.N.: *7426*

Date: *OCT 18 1995*



- a) Output signal amplitude:  
*TTL LEVELS*
- b) Pulse width:  
*5 ns TO 5 us*
- c) DIVISION  
FRONT: *2 TO 255*
- d) INPUT  
AMP: *0.1 TO 5 VOLTS*
- e) PRF: *1 Hz TO 100 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:  
*120 / 240 V RMS*  
*57 - 60 Hz*

INPUT: *50 MHz SINEWAVE*  
*.2 V RMS (P-P)*

*[Signature]*