


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ACMM-1-C-P-PN*  
S.N.: *7440*  
Date: *OCT 24 1995*



- a) Output signal amplitude:  
*0 TO ±2 VOLTS*
- b) Pulse width:  
*1 TO 10 NS*
- c) Rise time:  
*≤ 0.3 NS*
- d) Fall time:  
*≤ 0.3 NS*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*O.K.*
- g) Prime power:

*120/240 VOLTS*  
*50-60 MHz*

*[Signature]*

*40 db ATTEN*  
*.5 V/DIV*  
*TOP: 5 NS/DIV*  
*(PRF ≈ 25 MHz)*

*BOT: 1 NS/DIV*  
*(PRF ≈ 25 MHz)*