


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVK-61-C-ATEW-PN-05-R5*

S.N.: *7441*

Date: *FEB 5 1986*



(A) *P_{out}, R_L = 100K*



(B) *N_{out}, R_L = 100K*

- a) Output signal amplitude:
0 TO ± 250 V (25%)
- b) Pulse width: *(R_L > 10K)*
100 NS TO 1 MS
- c) Rise time: *(20% MAX DUTY CYCLE)*
≤ 10 NS
- d) Fall time: *← GND*
≤ 20 NS
- e) PRF: *0 TO 1 KHz*
(20% MAX DUTY CYCLE)
- f) Jitter, stability:
6%
- g) Prime power: *← GND*
120 / 280 V_{OUT}
50 - 60 Hz.

[Signature]