


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-3-C-PDN-0T*

S.N.: *7446*

Date: *DEC 20 1995*



- a) Output signal amplitude:
50-100 MHz: 0 TO ±15V
150 MHz: 0 TO ±10V
- b) Pulse width: *250 MHz: ±5.0V*
0.3 TO 1.0N)
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 150 PS
- e) PRF: *50 TO 250 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

100 ± 10%
(20 ± 10%)
50-60 MHz

40db ATTN
10 VOLTS / DIV
PRF = 100 MHz
TOP 1 NS / DIV
BOI 2 NS / DIV

[Signature]