


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-HVI-PN-SSC1

S.N.: 7463

Date: OCT. 13, 1995



- a) Output signal amplitude:  
 $\pm 100$  VOLTS
- b) Pulse width:  
0.6 TO 2.0 NS
- c) Rise time:  
N/A
- d) Fall time:  
N/A
- e) PRF:  
0 TO 100 kHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLTS  
250 mA MAX

TOP: P OUT

BOT: N OUT

60 dB ATTEN

100 VOLTS/DIV

PRF = 30 kHz