


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HVI-PN-SSC1

S.N.: 7465

Date: OCT 16 1995



- a) Output signal amplitude:
± 100 VOLTS -
- b) Pulse width:
0.6 TO 2.0 NS ✓
- c) Rise time:
N/A -
- d) Fall time:
N/A -
- e) PRF:
0 TO 100 kHz ✓
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS
250 mA MAX

TOP: P OUT
BOT: N OUT
60 dB ATTEN
100 VOLTS/DIV
PRF = 30 kHz