


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-4V1-PS-SSC1*

S.N.: *7467*

Date: *OCT 20 1995*



- a) Output signal amplitude: *± 100 V RMS*
- b) Pulse width: *0.6 TO 2.0 NS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:

TDP: P_{OUT}
Bot: N_{OUT}
60 db ATTN
100 V RMS / DIV
10 MF = 30 KHz

413 V RMS
250 mA
100 W

[Signature]