


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-141-PN-554

S.N.: 7460

Date: OCT 20 1995



- a) Output signal amplitude:  $\pm 100$  VOLTS
- b) Pulse width: 0.6 TO 2.0 NS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 100 KHz
- f) Jitter, stability: OK
- g) Prime power:  $\pm 15$  VOLTS

250 mA  
MAX.

*[Signature]*

TOP: P<sub>out</sub>

BOT: N<sub>out</sub>

60 db ATTEN

100 VOLTS/DIV

PRF = 30 KHz