


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-421-PN-SSC1

S.N.: 7469

Date: NOV 3 1995



- a) Output signal amplitude:  
 $\pm 100 \text{ V}$
- b) Pulse width:  
 $0.6 \text{ TO } 2.0 \text{ ns}$
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
 $0 \text{ TO } 100 \text{ kHz}$
- f) Jitter, stability:  
OK
- g) Prime power:

+15 volts  
250mA (MAX)  
L

TOP: Pout  
BOT: Nout  
60 Ohm ATTEN  
100 VOLT/DIV  
PRF = 30 kHz