


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MH-MVI-PA-55C*  
 S.N.: *7470*  
 Date: *NOV 3 1995*



- a) Output signal amplitude: *±100V*
- b) Pulse width: *0.6 TO 2.0nS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *±15 VDC*  
*250 mA*

*TOP: Pout*  
*BOT: Nout*

*60 Ohm Load*  
*100 VOLT/DIV*  
*VARF = 70KHz*

*MAX*  
*[Signature]*