


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVN-2-P

S.N.: 7475

Date: FEB 12 1996



a) Output signal amplitude:

0 to +2 V

b) Pulse width:

0.2 to 1.5 ns

c) Rise time:

$\leq 100$  ps

d) Fall time:

$\leq 135$  ps

e) PRF:

50 to 250 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 VDC

500 mA (max)

40 dB ATTEN

5 V0 203 / DIU.