


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-P.

S.N.: 7492

Date: Feb 26 1996



- a) Output signal amplitude:
0 TO +40 V_{OUT}
- b) Pulse width: $R_c = 500$
2 TO 100 ns
- c) Rise time:
 ≤ 350 ns
- d) Fall time:
 ≤ 2 ns.
- e) PRF: MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
200 mA.

50 dB ATTEN

32 VOLTS/DIV