


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-PW-W1-05-PWT-P*

S.N.: *7505*

Date: *FEB 06, 1996*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: *($R_L \geq 50\Omega$)*
50NS TO 1.0U
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
 $\leq 10NS$
- d) Fall time:
 $\leq 10NS$
- e) PRF: *0 TO 100 KHZ*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:

+24 VDC

500mA (MAX)

[Signature]

$R_L = 50\Omega$
PRF = 5 KHZ.