


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP-M-1-C-M-OT-N*

S.N.: *7510*

Date: *JUN 29 1996*



- a) Output signal amplitude:  
*0 TO -10 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 VOLTS*  
*50-60 HZ.*

*40 dB ATTEN*  
*-10 VOLTS/DIV*  
*BAF ≈ 300 MHz*