


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5A1-C-N-T3*

S.N.: *7520*

Date: *MAR 5 1986*




- a) Output signal amplitude: *0 TO -13 AMP (MAX)*
- b) Pulse width: *5 TO 100 NS*
- c) Rise time: *≤ 0.6 NS*
- d) Fall time: *≤ 7 NS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*

(A)

*-T. OUT TO. ≈ 10 Ω
60 db ATTEN. ∴ 100 VOLTS/DIV
TOP: 10 NS/DIV
DOT: 1 NS/DIV
PRF: 3 KHz.*

g) Prime power: *120/240 Volts, 50-60 Hz.*



MAIN FRAME OUT TO 50 Ω (60 db ATTEN)