


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-M-1-C-05-N-PN*

S.N.: *7528*

Date: *DEC 11 1995*



- a) Output signal amplitude:
0 TO 710 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN

10 VOLTS/DIV

TOP 1 NS/DIV

BOT 100 PS/DIV

PRF ≈ 100 KHz

120/240 V

500-60 MHz

[Signature]