


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVPP-1-C-CA-EN-EO-PN*

S.N.: *7529*

Date: *NOV 27 1995*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
A) 0.2 TO 5.0 NS
B) 5.0 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/DIV

120/240 VOLTS
50-60 MHz

TOP A RANGE:
1 NS/DIV

BOOT B RANGE:
5 NS/DIV

PRF ≈ 100 KHz