


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-P

S.N.: 7548

Date: JAN 16/96



- a) Output signal amplitude:
0 To + 30 V
- b) Pulse width:
 $\leq 1.0 \text{ NS}$
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 To 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VDC
200 mA (max)

40 db ATTEN

\therefore 10 VOLTS / DIV

PRF = 300 kHz