


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-102-P

S.N.: 7549

Date: JAN 16 1996



- a) Output signal amplitude:
7 + 60 V
- b) Pulse width:
≤ 2 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+135 VDC
5 mA (MAX)

[Signature]

40 dB ATTEN
10 VOLTS/DIV
PRF ≈ 300 kHz